



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory*, using *Bio-Rad Variant II* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>Variant II</b>	Calibrator Lot: <b>AN10566NU, AN10567NU</b>	Column Lot: <b>NU92450F</b>
Reagent Lot: <b>AA11093NU, AA11094NU, AA11095NU</b>	Calibrator Assigned Values: <b>5.9%, 9.9%</b>	

Date of Certification: January 1, 2012

Certification Expires: January 1, 2013

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor