




Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Gen Lab**, using **Roche Cobas c513 Dried Blood Spot ADX 100** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Roche Cobas c513 DBS ADX 100	Calibrator Lot: 740394	Application: HbA_{1c}% = (HbA_{1c}/Hb) x 91.5 + 2.15
Reagent Lot: 801755, 809704	Calibrator Assigned Values: 0.225, 0.56, 1.095, 1.51, 2.083, 2.959 g/dL HbA_{1c}; 2.4, 12.91 g/dL Hb	

Date of Certification: July 1, 2025

Certification Expires: July 1, 2026

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Curt Rohlfing".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read "Rhonda Haward".

SRL director/ supervisor