



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Roche Cobas c513 with DBS on ADX 100* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: *University of Missouri SRL#12*

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 671824	Application: HbA_{1c}%=(HbA_{1c}/Hb) x 91.5+2.15 DBS (ADX 100)
Reagent Lot: 75130401, 758478	Calibrator Assigned Values: 0.253, 0.584, 1.113, 1.526, 2.095, 2.962 g/dL HbA_{1c}; 2.38, 12.83 g/dL Hb	

Date of Certification: July 1, 2024

Certification Expires: July 1, 2025

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor