



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Gen Lab**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: <b>Cobas c513</b>	Calibrator Lot: <b>671824</b>	Application: <b>HbA1c%=(HbA1c/Hb)x91.5+2.15</b>
Reagent Lot: <b>715055, 746345</b>	Calibrator Assigned Values: <b>0.253, 0.584, 1.113, 1.526, 2.095, 2.962 g/dL HbA1c; 2.38, 12.83 g/dL Hb</b>	

Date of Certification: June 1, 2024

Certification Expires: June 1, 2025

*Daniel Sacks*

NGSP Steering Committee Chair

*Randie R Little PhD*

NGSP Network Coordinator

*Sharon Amalf*

SRL director/ supervisor