



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Roche Cobas c513* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: Roche Cobas c513	Calibrator Lot: 740394	Application: HbA1c% = (HbA1c/Hb) x 91.5 + 2.15
Reagent Lot: 899283, 911517, 88906001	Calibrator Assigned Values: 0.225, 0.560, 1.095, 1.510, 2.083, 2.959 g/dL HbA1c; 2.4, 12.91 g/dL Hb	

Date of Certification: June 1, 2026

Certification Expires: June 1, 2027

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor