



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Roche Cobas c513* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 728053	Application: HbA1c (%)=(HbA1c/Hb) *91.5+2.15
Reagent Lot: 804154, 829966, 817147	Calibrator Assigned Values: 0.254, 0.591, 1.124, 1.541, 2.119, 2.993 mmol/L HbA1c; 2.42, 13.01 mmol/L Hb	

Date of Certification: June 1, 2025

Certification Expires: June 1, 2026

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor