




Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory - Olathe**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Roche Cobas c513	Calibrator Lot: 800811	Application: HbA1c% = (HbA1c/Hb) x 91.5 + 2.15
Reagent Lot: 886047, 886781	Calibrator Assigned Values: 0.246, 0.544, 1.021, 1.391, 1.905, 2.684 g/dL HbA1c; 2.38, 12.82 g/dL Hb	

Date of Certification: September 1, 2026

Certification Expires: September 1, 2027

A handwritten signature in black ink, appearing to read "David Sacks".

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Curt Rohlfing".

NGSP Network Coordinator

A handwritten signature in black ink, appearing to be a stylized monogram.

SRL director/ supervisor