



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Olathe**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: <b>Cobas c513</b>	Calibrator Lot: <b>626346</b>	Application: <b>HbA1c%=(HbA1c/Hb)x91.5+2.15</b>
Reagent Lot: <b>715055, 718315, 756498</b>	Calibrator Assigned Values: <b>0.240, 0.564, 1.077, 1.475, 2.029, 2.871 g/dL HbA1c; 2.33, 12.53 g/dL Hb</b>	

Date of Certification: September 1, 2024

Certification Expires: September 1, 2025

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor