


# Certificate of Traceability



## Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Gen Lab*, using *Roche Cobas c513 DBS Whatman* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: *University of Missouri SRL#12*

The system evaluated was:

Instrument: <b>Cobas c513 DBS Whatman</b>	Calibrator Lot: <b>671824</b>	Application: <b>HbA1c%=(HbA1c/Hb)x 91.5+2.15</b>
Reagent Lot: <b>751304, 758478</b>	Calibrator Assigned Values: <b>0.253, 0.584, 1.113, 1.526, 2.095, 2.962 g/dL HbA1c; 2.38, 12.83 g/dL Hb</b>	

Date of Certification: August 1, 2024

Certification Expires: August 1, 2025

Handwritten signature of David Sacks.

NGSP Steering Committee Chair

Handwritten signature of Randie R Little PhD.

NGSP Network Coordinator

Handwritten signature of Sharon Anulf.

SRL director/ supervisor