




Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Gen Lab**, using **Sebia CAPILLARYS 3 Tera** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Sebia CAPILLARYS 3 Tera	Calibrator Lot: 23033, 24033	Application: HbA1c% = HbA1c/(HbA1c + HbA0)
Reagent Lot: 14055, 20035, 11065	Calibrator Assigned Values: 35.5 mmol/mol, 83.7 mmol/mol	

Date of Certification: February 1, 2026

Certification Expires: February 1, 2027

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor