




Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory - Olathe**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: Roche Cobas c513	Calibrator Lot: 740394	Application: HbA1c% = (HbA1c/Hb) x 91.5 =2.15
Reagent Lot: 801755, 809704	Calibrator Assigned Values: 0.219, 0.551, 1.080, 1.492, 2.059, 2.925 g/dL HbA1c; 2.40, 12.91 g/dL Hb	

Date of Certification: September 1, 2025

Certification Expires: September 1, 2026

A handwritten signature in black ink, appearing to read 'David Sacks'.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Curt Rohlfing'.

NGSP Network Coordinator

A handwritten signature in black ink, appearing to read 'Rhonda Haward'.

SRL director/ supervisor