

CHOLESTEROL REFERENCE METHOD LABORATORY NETWORK

CERTIFICATE OF TRACEABILITY

This certifies that

*Clinical Reference Laboratory
Lenexa, KS, USA*

has documented traceability to the National Reference System for Cholesterol by performing a direct comparison with the cholesterol reference method using fresh patient specimens. This laboratory has demonstrated the ability to meet the National Cholesterol Education Program's performance criteria for accuracy. The comparison was performed with

University of Washington, Department of Medicine
Northwest Lipid Metabolism and Diabetes Research Laboratories
Seattle, Washington

The system evaluated was:

Instrument:

*Roche Diagnostics
Cobas 8000 1138-04*

Reagent:

*Roche Diagnostics Chol Reagent
Lot #: 23317501*

Calibrator:

*Roche Diagnostics c.f.a.s.
Lot #: 19883001*

Sample Matrix:

Venous Serum

Date of evaluation: 07/14/2017

Certificate expires: 01/14/2018

CRMLN Laboratory Director: _____

[Signature]