



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory – Gen Lab**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 16006101	Application: HbA1c%=(HbA1c/Hb)x
Reagent Lot: 25528001, 28326001, 30489401	Calibrator Assigned Values: 0.230, 0.584, 1.151, 1.589, 2.198, 3.123 g/dL HbA1c; 2.29, 12.27 g/dL Hb	91.5+2.15

Date of Certification: June 1, 2018

Certification Expires: June 1, 2019

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor