



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory - EU**, using **Roche Cobas c501** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: <b>Cobas c501</b>	Calibrator Lot: <b>16006100</b>	Application: <b>HbA1c%=(A1-H3/HB-H3)x91.5+2.15</b>
Reagent Lot: <b>774553, 269182</b>	Calibrator Assigned Values: <b>0.222, 0.568, 1.13, 1.57, 2.148, 3.067 g/dL HbA1c; 2.178, 11.743 g/dL Hb</b>	

Date of Certification: February 1, 2018

Certification Expires: February 1, 2019

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor