



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Gen Lab**, using **Roche Cobas c513 Dried Blood Spot** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 16006600	Application: HbA_{1c}%=(HbA_{1c}/Hb)x91.5 +2.15
Reagent Lot: 25528001, 28262201, 31880201	Calibrator Assigned Values: 0.230, 0.584, 1.151, 1.589, 2.198, 3.123 g/dL HbA_{1c}; 2.29, 12.27 g/dl Hb	

Date of Certification: July 1, 2018

Certification Expires: July 1, 2019

Daniel Sarkis

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor