



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory-Gen Lab**, using **Roche Cobas c513 Dried Blood Spot** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 257104	Application: HbA1c%=(HbA1c/Hb)x91.5+2.15 Dried Blood Spot
Reagent Lot: 341664, 32774601	Calibrator Assigned Values: 0.219, 0.551, 1.080, 1.492, 2.062, 2.932 g/dL HbA1c; 2.40, 12.94 g/dL Hb	

Date of Certification: July 1, 2019

Certification Expires: July 1, 2020

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalfi

SRL director/ supervisor