



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory – Gen Lab**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: <b>Cobas c513</b>	Calibrator Lot: <b>257104</b>	Application: <b>HbA1c%=(HbA1c/Hb)x91.5+2.15</b> <b>Whole Blood</b>
Reagent Lot: <b>341664, 39216001</b>	Calibrator Assigned Values: <b>0.219, 0.551, 1.080, 1.492, 2.062, 2.932 g/dL</b> <b>HbA1c; 2.40, 12.94 g/dL Hb</b>	

Date of Certification: June 1, 2019

Certification Expires: June 1, 2020

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor