



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory - EU**, using **Roche Cobas C501** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas C501	Calibrator Lot: 229753	Application: HbA1c% = (A1-H3/HB-H3) x 91.5 + 2.15
Reagent Lot: 34484501, 32774601	Calibrator Assigned Values: 0.211, 0.539, 1.071, 1.492, 2.038, 2.912 g/dL HbA1c; 2.289, 12.342 g/dL Hb	

Date of Certification: February 1, 2019

Certification Expires: February 1, 2020

Daniel Sacks
NGSP Steering Committee Chair

Randie R Little PhD
NGSP Network Coordinator

Sharon Amalfi
SRL director/ supervisor