



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Clinical Reference Laboratory - Olathe**, using **Roche Cobas c513** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 160066	Column Lot: HbA1c%=(HbA1c/Hb)x91.5+2.15 Gen.3
Reagent Lot: 211965, 283260, 304894	Calibrator Assigned Values: 0.242, 0.583, 1.125, 1.549, 2.133, 3.024 g/dL HbA1c; 2.29, 12.27 g/dL Hb	

Date of Certification: September 1, 2018

Certification Expires: September 1, 2019

Daniel Sachs
NGSP Steering Committee Chair

Randie R Little PhD.
NGSP Network Coordinator

Shawn Amalf
SRL director/ supervisor