



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Clinical Reference Laboratory-Olathe*, using *Roche Cobas c513* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Cobas c513	Calibrator Lot: 257104	Application: HbA1c%=(HbA1c/Hb)x
Reagent Lot: 341664, 32712001, 38142101	Calibrator Assigned Values: 0.219, 0.551, 1.080, 1.492, 2.062, 2.932 g/dL HbA1c; 2.40, 12.94 g/dL Hb	91.5+2.15 Gen.3

Date of Certification: September 1, 2019

Certification Expires: September 1, 2020

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor